



APPENDIX "E"

TYPICAL INSTRUMENT DETAILS

Texas Instruments Model 10,000 Amplifier System

(a) Recording Settings

Initial gain	40-43		
Final gain	90-95		
GAGC expansion	80 db/sec		
AGC level	6-10		
Filters		frequency	slope
	High cut	70-92	24
	Low cut	8	12

(b) Playbacks - direct

Texas Instruments 10 K S. D. F. S.

Channel 1-24	Seismic Trace 1-24
25	
26	WB 5
27	WB 20
28	WB 24
29	WB 10
30	Gain
31	WB 1, 15
0	Timing word